

## Material Information Bio Herador CN

High gold content ceramic bonding alloy  
(according to EN ISO 22674 und EN ISO 9693-1)

### Chemical Composition (Content in mass %):

Bio Herador CN	Au	Pt	Zn	In	Ir	Mn	Ta	Rh
	86,6	10,4	1,5	0,2	0,1	0,1	0,2	0,9

This Dental Alloy is free of Beryllium, Cadmium and Lead.

**Manufacturer:** Kulzer GmbH

**Bio-compatibility:** We certify that this product has been evaluated according to the internationally valid directives EN ISO 7405, "Dentistry – Evaluation of biocompatibility of medical devices used in dentistry" and EN ISO 10993 series, "Biological evaluation of medical devices". The evaluation covers, among other things, possible risk of cytotoxicity, allergisation, irritation, and genotoxicity. The studies conducted were carried out in independent testing facilities according to OECD directives and fulfil the requirements of GLP (Good Laboratory Practice). The evaluation proves the biological compatibility of this dental alloy.

**Corrosion:** Corrosion testing was carried out at Kulzer's testing facility according to the EN ISO 22674: „Dentistry - Metallic materials for fixed and removable restorations and appliances". The directive stipulates a total release of metal ions from the metallic dental material of not more than 200 µg/cm<sup>2</sup> over 7 days at 37°C. The tested alloy showed a metal ion release less than 10 µg/cm<sup>2</sup>. This proves, it is a high corrosion resistant alloy.

**Evaluation:** Based on the above test and evaluation, this alloy is classified as biocompatible (biologically compatible) when used as intended. A clinical evaluation proved that specified characteristics and performance related requirements are fulfilled.

**Ceramic-compatibility:** According to the instruction for use, the dental alloy Bio Herador CN, with a CTE (25°C – 500°C) of 14,5 µm/mK can be considered as compatible to HeraCeram.

**The conformity of Bio Herador CN with internationally valid directives is proven and certified by the Technical Control Board of Rheinland (CE 0197). This document is subject to an internal approval process. It was created electronically and is valid without signature.**

### Contact in Germany

Kulzer GmbH  
Leipziger Straße 2  
63450 Hanau, Germany  
info.lab@kulzer-dental.com  
kulzer.com

Giving a hand to oral health.



**KULZER**  
MITSUI CHEMICALS GROUP